

Corrigendum I dated 12-06-2025 for SHWETDHARA: AYODHYA: SITC of Various Milk Testing Component/2025-26/01

Pre bid meeting held with Shwethdhara on 07 June 2025 at 11.00 HRS. The reply of Purchaser given as under

Following were attended the meeting:

- 1- Prompt
- 2- Kanha
- 3- Everest
- 4- Pooransree Electronics
- 5- Reil
- 6- Qboid
- 7- STIPL
- 8- Samudra Technologies

| Sr No. | Bidder | Queries | Reply from Purchaser |
|---------------|---------------|---|--|
| 1 | Prompt | how to ensure wisense compatibility interact, sample of wisense should be given before bid submission, trial demonstration must be taken to understand the interface with wisens device | To plug the Wi-sense directly on any device (MA, EWS, DI), equipment should have following connection. 1. 9 Pin Male RS232 Male Connector 2. 5V Power Supply on #9 number pin. 3. 2# number pin should be Rx 4. #3 number pin should be Tx 5. #5 number pin should be ground. |
| 2 | Prompt | kindly give Radius Area of service from your plant may be so that we can plan our manpower efficiently to serve you better | 50-70 KM |
| 3 | Prompt | after installation of 30 days penalty clause should be liberalised to understand the operational training etc. | As per tender document |

| | | | |
|---|--------|--|--|
| 4 | Prompt | <p>on the portal there is solar panel to quote but it contains battery also which takes 28% GST while Solar Panel takes 12% GST and there is option to select one slab of GST only, how to quote please suggest</p> | <p>Please note that it is recommended for the supplier to indicate a single applicable GST rate, as in many cases items may be procured on a composite basis, and a uniform GST rate would apply.</p> <p>It should also be noted that the price evaluation of each item will be based on the total price inclusive of GST, irrespective of the GST rate quoted by the supplier. No price increase will be entertained after bid submission, and if the bidder refuses to honour their quoted price after the price bid is opened, their EMD may be forfeited.</p> <p>For example, if a supplier quotes a unit price of ₹100 with an applicable GST of 18%, the total price including GST would be ₹118. However, if the actual applicable GST as per the HSN code is 28%, the purchaser will reverse-calculate the base price from ₹118. In this case, the basic price would be ₹92.1875, and 28% GST on that amount (₹25.8125) would again result in ₹118.</p> <p>Hence, the price including GST (₹118 in this example) will be considered for evaluation purposes. Bidders are advised to quote accordingly.</p> |
| 1 | KANHA | <p>We request that the payment terms be aligned with the previous tender (Indujaa), allowing for milestone-based or partial payments instead of 100% post-installation payment. This will ease working capital burden and ensure smoother execution.</p> | <p>Revised payment term as under : 60% payment shall be made within 30 days after Completion of supply and against the submission of BG of 10% amount of contract Value including GST. Balance 40% shall be made after installation and commission of equipment at site. All the bank guarantee/s should be in the pro-forma prescribed by purchaser and from a Nationalized Bank or a foreign bank having branches in India. The performance/retention bank guarantees should be valid up to 90 days beyond the warranty period.</p> |
| 2 | KANHA | <p>Kindly consider evaluation on an overall basis instead of item-wise evaluation as mentioned. A 60:40 bifurcation can be maintained for L1 and L2 bidders, but an</p> | <p>No. As per tender document</p> |

| | | | |
|---|-------|--|--|
| | | overall package evaluation would ensure better price optimization and vendor efficiency. | |
| 3 | KANHA | Please share an indicative radius or geographical scope (district-wise/block-wise) of the installation locations under Shwethdara MPCL. This will help us estimate the service engineer costing accurately as 48-hour service response time is mandatory. | 50-70 KM |
| 4 | KANHA | There is ambiguity in EMD submission – one clause specifies RTGS/NEFT while another mentions a format for DD. Kindly confirm the acceptable payment mode and clarify if both options are valid or only RTGS/NEFT is applicable. | EMD will not be acceptable in the form of DD, only through RTGS/NEFT mode |
| 1 | REIL | Please confirm the applicable Payment Terms & we suggest payment terms should be as follows:-(i) 80% payment should be made against delivery of material.(ii) Remaining 20% Payment should be made within 90 days of delivery of the material. We propose that payment against I & C shall be in closed tenure/period. | Revised payment term as under : 60% payment shall be made within 30 days after Completion of supply and against the submission of BG of 10% amount of contract Value including GST. Balance 40% shall be made after installation and commission of equipment at site. All the bank guarantee/s should be in the pro-forma prescribed by purchaser and from a Nationalized Bank or a foreign bank having branches in India. The performance/retention bank guarantees should be valid up to 90 days beyond the warranty period. |
| 2 | REIL | All the equipment to be supplied should be compatible with the connecting device (Wi-Sense) of | To plug the Wi-sense directly on any device (MA, EWS, DI), equipment should have following connection. 1. 9 Pin Male RS232 Male Connector 2. 5V Power Supply on #9 number pin. |

| | | | |
|---|-------|---|--|
| | | Shwethdara Kindly clear the following: Scope of protocols with Complete SOPs of App . | 3. 2# number pin should be Rx 4. #3 number pin should be Tx 5. #5 number pin should be ground. |
| 1 | STIPL | Kindly confirm the exact specifications, make, and model of the Wi-Sense device to be used for compatibility assessment. | To plug the Wi-sense directly on any device (MA, EWS, DI), equipment should have following connection. 1. 9 Pin Male RS232 Male Connector 2. 5V Power Supply on #9 number pin. 3. 2# number pin should be Rx 4. #3 number pin should be Tx 5. #5 number pin should be ground. |
| 2 | STIPL | Bank Guarantee (BG) Submission Please confirm whether the BG can be submitted through Online Transfer/NEFT/RTGS or only via traditional Bank Guarantee format. | 1. Performance BG Acceptable in BG format given in the tender where as supplier can also make the payment for the same amount by NEFT/RTGS . 2. Whereas EMD only via RTGS/NEFT mode |
| 3 | STIPL | It may be amended as – “System having design of Single or Double Sensor.” Kindly remove the phrase “and which should be mounted on socket” to allow flexibility in design. | No. As per specification of tender document |
| 4 | STIPL | Please allow the specification to read as – <i>“AISI 304/SUS 304 (with certificate)”</i> for broader material equivalence. | No. As per specification of tender document |
| 5 | STIPL | Penalty Clause We request that the penalty clause be made applicable only after the complete installation of all systems and the associated training is successfully completed, rather than from the date of initial installation. | No. As per specification of tender document |

| | | | |
|---|-------|---|----------|
| 6 | STIPL | Service Engineer Deployment Radius Kindly share the expected radius of area of operation where service engineers are required to be deputed, so that we can plan accordingly. | 50-70 KM |
| | | Last date of Submission of offer is 16-06-2025 | |