

Corrigendum I dated 12th June 2025 for MARATHVARHAD:NAGPUR: SITC of Various Milk Testing Component/2025-26/01

Pre bid meeting held with MARATHVARHAD MPO on 09th June 2025 at 15.00 HRS. The reply of Purchaser given as under

Pre-bid Meeting is attended by the following Bidder: Kanha Milk Testing Equipment's Pvt. Ltd., Everest Instrument Pvt. Ltd. and Prompt Equipments Pvt. Ltd.

Sr No.	Queries	Reply from Purchaser																								
1	how to ensure wisense compatibility interact, sample of wisense should be given before bid submission, trial demonstration must be taken to understand the interface with wisens device	To plug the Wi-sense directly on any device (MA, EWS, DI), equipment should have following connection. 1. 9 Pin Male RS232 Male Connector 2. 5V Power Supply on #9 number pin. 3. 2# number pin should be Rx 4. #3 number pin should be Tx 5. #5 number pin should be ground.																								
2	kindly give Radius Area of service from your plant may be so that we can plan our manpower efficiently to serve you better	Services shall be provided by successful bidder in the 100 KM radius of below delivery locations as per below delivery schedule. <table border="1" data-bbox="974 1027 2011 1409"> <thead> <tr> <th data-bbox="974 1027 1570 1141">Delivery Location in Vidarbha & Marathwada in Maharashtra</th> <th data-bbox="1570 1027 1667 1141">UOM</th> <th data-bbox="1667 1027 2011 1141">Qty. of Milk Testing Components</th> </tr> </thead> <tbody> <tr> <td data-bbox="974 1141 1570 1182">Malkapur Dist. Buldana</td> <td data-bbox="1570 1141 1667 1182">Nos.</td> <td data-bbox="1667 1141 2011 1182">45</td> </tr> <tr> <td data-bbox="974 1182 1570 1222">Karla, Dist. Amrawati</td> <td data-bbox="1570 1182 1667 1222">Nos.</td> <td data-bbox="1667 1182 2011 1222">45</td> </tr> <tr> <td data-bbox="974 1222 1570 1263">Nagpur, Dist. Nagpur</td> <td data-bbox="1570 1222 1667 1263">Nos.</td> <td data-bbox="1667 1222 2011 1263">125</td> </tr> <tr> <td data-bbox="974 1263 1570 1304">DHANOLI, Dist. Wardha</td> <td data-bbox="1570 1263 1667 1304">Nos.</td> <td data-bbox="1667 1263 2011 1304">45</td> </tr> <tr> <td data-bbox="974 1304 1570 1344">Laur Dist.- Latur</td> <td data-bbox="1570 1304 1667 1344">Nos.</td> <td data-bbox="1667 1304 2011 1344">45</td> </tr> <tr> <td data-bbox="974 1344 1570 1385">Badnapur, Dist: Jalna</td> <td data-bbox="1570 1344 1667 1385">Nos.</td> <td data-bbox="1667 1344 2011 1385">45</td> </tr> <tr> <td data-bbox="974 1385 1570 1409">Bhoom, Dist.- Dharashiv</td> <td data-bbox="1570 1385 1667 1409">Nos.</td> <td data-bbox="1667 1385 2011 1409">50</td> </tr> </tbody> </table>	Delivery Location in Vidarbha & Marathwada in Maharashtra	UOM	Qty. of Milk Testing Components	Malkapur Dist. Buldana	Nos.	45	Karla, Dist. Amrawati	Nos.	45	Nagpur, Dist. Nagpur	Nos.	125	DHANOLI, Dist. Wardha	Nos.	45	Laur Dist.- Latur	Nos.	45	Badnapur, Dist: Jalna	Nos.	45	Bhoom, Dist.- Dharashiv	Nos.	50
Delivery Location in Vidarbha & Marathwada in Maharashtra	UOM	Qty. of Milk Testing Components																								
Malkapur Dist. Buldana	Nos.	45																								
Karla, Dist. Amrawati	Nos.	45																								
Nagpur, Dist. Nagpur	Nos.	125																								
DHANOLI, Dist. Wardha	Nos.	45																								
Laur Dist.- Latur	Nos.	45																								
Badnapur, Dist: Jalna	Nos.	45																								
Bhoom, Dist.- Dharashiv	Nos.	50																								

		Total	400
3	after installation of 30 days penalty clause should be liberalised to understand the operational training etc.	As per tender document	
4	on the portal there is solar panel to quote but it contains battery also which takes 28% GST while Solar Panel takes 12% GST and there is option to select one slab of GST only, how to quote please suggest	<p>Please note that it is recommended for the supplier to indicate a single applicable GST rate, as in many cases items may be procured on a composite basis, and a uniform GST rate would apply.</p> <p>It should also be noted that the price evaluation of each item will be based on the total price inclusive of GST, irrespective of the GST rate quoted by the supplier. No price increase will be entertained after bid submission, and if the bidder refuses to honour their quoted price after the price bid is opened, their EMD may be forfeited.</p> <p>For example, if a supplier quotes a unit price of ₹100 with an applicable GST of 18%, the total price including GST would be ₹118. However, if the actual applicable GST as per the HSN code is 28%, the purchaser will reverse-calculate the base price from ₹118. In this case, the basic price would be ₹92.1875, and 28% GST on that amount (₹25.8125) would again result in ₹118.</p> <p>Hence, the price including GST (₹118 in this example) will be considered for evaluation purposes. Bidders are advised to quote accordingly.</p>	

5	<p>We request that the payment terms be aligned with the previous tender (Indujaa), allowing for milestone-based or partial payments instead of 100% post-installation payment. This will ease working capital burden and ensure smoother execution.</p>	<p>Revised payment term as under : 60% payment shall be made within 30 days after Completion of supply and against the submission of BG of 10% amount of contract Value including GST. Balance 40% shall be made after installation and commission of equipment at site. All the bank guarantee/s should be in the pro-forma prescribed by purchaser and from a Nationalized Bank or a foreign bank having branches in India. The performance/retention bank guarantees should be valid up to 90 days beyond the warranty period.</p>																											
6	<p>Kindly consider evaluation on an overall basis instead of item-wise evaluation as mentioned. A 60:40 bifurcation can be maintained for L1 and L2 bidders, but an overall package evaluation would ensure better price optimization and vendor efficiency.</p>	<p>No. As per tender document</p>																											
7	<p>Please share an indicative radius or geographical scope (district-wise/block-wise) of the installation locations under MARATHVARHAD MPCL. This will help us estimate the service engineer costing accurately as 48-hour service response time is mandatory.</p>	<p>Services shall be provided by successful bidder in the 100 KM radius of below delivery locations as per below delivery schedule.</p> <table border="1" data-bbox="978 997 2022 1427"> <thead> <tr> <th data-bbox="978 997 1602 1114">Delivery Location in Vidarbha & Marathwada in Maharashtra</th> <th data-bbox="1602 997 1787 1114">UOM</th> <th data-bbox="1787 997 2022 1114">Qty. of Milk Testing Components</th> </tr> </thead> <tbody> <tr> <td data-bbox="978 1114 1602 1154">Malkapur Dist. Buldana</td> <td data-bbox="1602 1114 1787 1154">Nos.</td> <td data-bbox="1787 1114 2022 1154">45</td> </tr> <tr> <td data-bbox="978 1154 1602 1195">Karla, Dist. Amrawati</td> <td data-bbox="1602 1154 1787 1195">Nos.</td> <td data-bbox="1787 1154 2022 1195">45</td> </tr> <tr> <td data-bbox="978 1195 1602 1235">Nagpur, Dist. Nagpur</td> <td data-bbox="1602 1195 1787 1235">Nos.</td> <td data-bbox="1787 1195 2022 1235">125</td> </tr> <tr> <td data-bbox="978 1235 1602 1276">DHANOLI, Dist. Wardha</td> <td data-bbox="1602 1235 1787 1276">Nos.</td> <td data-bbox="1787 1235 2022 1276">45</td> </tr> <tr> <td data-bbox="978 1276 1602 1317">Laur Dist.- Latur</td> <td data-bbox="1602 1276 1787 1317">Nos.</td> <td data-bbox="1787 1276 2022 1317">45</td> </tr> <tr> <td data-bbox="978 1317 1602 1357">Badnapur, Dist: Jalna</td> <td data-bbox="1602 1317 1787 1357">Nos.</td> <td data-bbox="1787 1317 2022 1357">45</td> </tr> <tr> <td data-bbox="978 1357 1602 1398">Bhoom, Dist.- Dharashiv</td> <td data-bbox="1602 1357 1787 1398">Nos.</td> <td data-bbox="1787 1357 2022 1398">50</td> </tr> <tr> <td colspan="2" data-bbox="978 1398 1787 1427">Total</td> <td data-bbox="1787 1398 2022 1427">400</td> </tr> </tbody> </table>	Delivery Location in Vidarbha & Marathwada in Maharashtra	UOM	Qty. of Milk Testing Components	Malkapur Dist. Buldana	Nos.	45	Karla, Dist. Amrawati	Nos.	45	Nagpur, Dist. Nagpur	Nos.	125	DHANOLI, Dist. Wardha	Nos.	45	Laur Dist.- Latur	Nos.	45	Badnapur, Dist: Jalna	Nos.	45	Bhoom, Dist.- Dharashiv	Nos.	50	Total		400
Delivery Location in Vidarbha & Marathwada in Maharashtra	UOM	Qty. of Milk Testing Components																											
Malkapur Dist. Buldana	Nos.	45																											
Karla, Dist. Amrawati	Nos.	45																											
Nagpur, Dist. Nagpur	Nos.	125																											
DHANOLI, Dist. Wardha	Nos.	45																											
Laur Dist.- Latur	Nos.	45																											
Badnapur, Dist: Jalna	Nos.	45																											
Bhoom, Dist.- Dharashiv	Nos.	50																											
Total		400																											

8	<p>There is ambiguity in EMD submission – one clause specifies RTGS/NEFT while another mentions a format for DD. Kindly confirm the acceptable payment mode and clarify if both options are valid or only RTGS/NEFT is applicable.</p>	<p>EMD will not be acceptable in the form of DD, only through RTGS/NEFT mode</p>
9	<p>Please confirm the applicable Payment Terms & we suggest payment terms should be as follows:-(i) 80% payment should be made against delivery of material.(ii) Remaining 20% Payment should be made within 90 days of delivery of the material. We propose that payment against I & C shall be in closed tenure/period.</p>	<p>Revised payment term as under :</p> <p>60% payment shall be made within 30 days after Completion of supply and against the submission of BG of 10% amount of contract Value including GST. Balance 40% shall be made after installation and commission of equipment at site.</p> <p>All the bank guarantee/s should be in the pro-forma prescribed by purchaser and from a Nationalized Bank or a foreign bank having branches in India. The performance/retention bank guarantees should be valid up to 90 days beyond the warranty period.</p>
10	<p>All the equipment to be supplied should be compatible with the connecting device (Wi-Sense) of Marathvarhad, Kindly clear the following: Scope of protocols with Complete SOPs of App .</p>	<p>To plug the Wi-sense directly on any device (MA, EWS, DI), equipment should have following connection.</p> <ol style="list-style-type: none"> 1. 9 Pin Male RS232 Male Connector 2. 5V Power Supply on #9 number pin. 3. 2# number pin should be Rx 4. #3 number pin should be Tx 5. #5 number pin should be ground.

11	Kindly confirm the exact specifications, make, and model of the Wi-Sense device to be used for compatibility assessment.	To plug the Wi-sense directly on any device (MA, EWS, DI), equipment should have following connection. 1. 9 Pin Male RS232 Male Connector 2. 5V Power Supply on #9 number pin. 3. 2# number pin should be Rx 4. #3 number pin should be Tx 5. #5 number pin should be ground.
12	Bank Guarantee (BG) Submission Please confirm whether the BG can be submitted through Online Transfer/NEFT/RTGS or only via traditional Bank Guarantee format.	1. Performance BG Acceptable in BG format given in the tender where as supplier can also make the payment for the same amount by NEFT/RTGS . 2. Whereas EMD only via RTGS/NEFT mode
13	It may be amended as – “System having design of Single or Double Sensor.” Kindly remove the phrase “and which should be mounted on socket” to allow flexibility in design.	No. As per specification of tender document
14	Please allow the specification to read as – <i>“AISI 304/SUS 304 (with certificate)”</i> for broader material equivalence.	No. As per specification of tender document
15	Penalty Clause We request that the penalty clause be made applicable only after the complete installation of all systems and the associated training is	No. As per specification of tender document

successfully completed, rather than from the date of initial installation.

16

Service Engineer Deployment Radius
Kindly share the expected radius of area of operation where service engineers are required to be deputed, so that we can plan accordingly.

Services shall be provided by successful bidder in the 100 KM radius of below delivery locations as per below delivery schedule.

Delivery Location in Vidarbha & Marathwada in Maharashtra	UOM	Qty. of Milk Testing Components
Malkapur Dist. Buldana	Nos.	45
Karla, Dist. Amrawati	Nos.	45
Nagpur, Dist. Nagpur	Nos.	125
DHANOLI, Dist. Wardha	Nos.	45
Laur Dist.- Latur	Nos.	45
Badnapur, Dist: Jalna	Nos.	45
Bhoom, Dist.- Dharashiv	Nos.	50
Total		400

17

These cluses shall be the part of tender document

2.25 Bid Prices

- a. The Bidder shall indicate on the Price Bid at NCDFI e-Portal the final unit prices (after considering discount, if any) of the goods it proposes to supply under the Contract. Bidders must submit a bid for the full quantity specified under technical specifications, failing which, such bids will not be considered for evaluation and comparison and will not be considered for award. The item wise quoted price should be **inclusive of applicable GST.**

- b. The item wise price of goods/services to be entered in the NCDFI e-portal shall be on FOR site basis **inclusive of applicable GST**, but including the charges for packing and forwarding, transportation, transit insurance and all other local costs incidental to delivery of goods/services to their final destination, storage cum erection insurance and safe custody at site.

2.26 Prices indicated on the Price Bid shall be entered in the following manner:

- a. The item wise price of goods to be supplied/installed and commissioned shall be on FOR site basis **inclusive of applicable GST**. The item wise price shall also include the charges for packing and forwarding, transportation, transit insurance and all other local costs incidental to delivery of the goods to their final destination, storage cum Erection insurance and safe custody at site.
- b. The Bidder's separation of price components in accordance with above will be solely for the purpose of facilitating the comparison of bids by the Purchaser and will not in any way limit the Purchaser's right to contract on any of the terms offered.

2.27 Fixed Price

Basic Price quoted by the Bidder shall be fixed during the bidder's performance of the Contract and not subject to variation

on any account. A bid submitted with an adjustable price quotation for such goods and services will be treated as non-responsive and rejected pursuant to Clause 24.

2.28 Price Break-up

Bidders shall furnish the price for the supply and for installation/ commissioning along with detailed cost break-up (item wise), which will be applicable for progressive payments. Items and works for which no break-up price is furnished by the bidder will not be paid for by the purchaser when supplied/executed and shall be deemed covered by the other break-up prices.

- a. However for evaluation purpose, prices quoted as indicated in Para 2.26 (a) above only will be considered.

Notwithstanding anything stated elsewhere in the bidding documents, irrespective of mode of the contracting with the successful bidder, the successful bidder will be liable for the payment of Indian Income Tax, surcharge on Income Tax and any other Corporate Tax, turnover tax etc. if attracted under the provisions of the law. The purchaser shall not bear any tax liability whatsoever irrespective of the mode of contracting

	Last date of Submission of offer is -----	No Change.
--	---	------------